

Equipment Details for Calibration/Repair

Equipment Type	Serial No.	Manufacturer	Model	Volume	Problem Description

FEMTO FIELD ENGINEER INITIALS _____

Prior to service of equipment that may be contaminated with biohazardous materials, potentially biohazardous materials or radioactivity; the user must first decontaminate it. Please complete the form below and return it to Femto Scientific with your instruments.

The decontamination procedure will include the following as appropriate:

- This equipment **has not** been used in a laboratory / location exposing it to biohazards or radioactive materials.
- This equipment **has** been used in a laboratory / location exposing it to biohazardous or radioactive materials. This equipment was appropriately decontaminated from biohazardous material with:
 - Autoclaving Ethylene Oxide Irradiation 10% bleach and water Other: _____
- This equipment was appropriately decontaminated & tested for radioactivity using method:
 - Wipe test with 3H, 14C Geiger Counter Other: _____

INSTITUTION/COMPANY _____ BLDG _____ RM _____

NAME (print) _____ PHONE # _____

SIGNATURE _____ DATE _____

Federal regulations require that all instruments sent for service be free of radioactivity and biohazards.